



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample  
Lot Number 24066046

Ref Number 53360  
Exp Date 2025-08-27  
Date of Manufacture 2024-02-27  
UDI# (01)00385480533607(11)240227(17)250827(10)24066046

Cal A	L/N	24058019	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.356	N/A
pCO2 mmHg		24.5	N/A

Cal B	L/N	24058020	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	24058022	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		51.0	N/A
pO2 mmHg		98.0	N/A

Reference	L/N	24058051	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

03/15/2024  
Date