OI-922 Page 40 of 66 Rev. BJ Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	LIFESCAN
Description:	TEST STRIPS, STATSTRIP TM GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT
Lifescan Purchase Order No.:	472579
Catalog Number:	42214
Nova Biomedical P/N	54311
One Touch P/N	02307501
Target Range (mg/dL) SD(mg/dL) /CV (%)	
$30-50 \leq 4 \text{ mg/dL}$	PASS FAIL
51-110 ≤ 6.0%	PASS FAIL
111-150 $\leq 5.0\%$	PASS FAIL
$151-250 \leq 5.0\%$	PASS FAIL
$251-400 \leq 5.0\%$	PASS/ FAIL
Accuracy Criteria 1: Total within $\pm 12\%$ of reference method $\geq 95\%$	PASS FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +15% of reference method ≥ 98%	+/- PASS/ FAIL
Qty:	24, 192
Lot No.	0324043309
Date of Mfg. (DOM):	2024/02/12 (YYYY/MM/DD)
Expiration Date:	2024/08 (YYYY/MM)
Unique Device Identifier (UDI):	$\begin{array}{c} (01)20385480422141 \\ (11)240212 \\ (17)260812 \\ (10)0324043309 \end{array}$

Mar Browne 2/23/ Quality Control Manager Date

Nova Biomedical Corporation 200 Prospect St Waltham, MA 02454