



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 24053011

Ref Number 54652
Exp Date 2025-08-22
Date of Manufacture 2024-02-22
UDI# (01)00385480546522(11)240222(17)250822(10)24053011

| Ctl Level 1 | | | Ctl Level 2 | | |
|-------------|-------------|------------------|-------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| HDL mg/dL | 28 - 50 | Pass | HDL mg/dL | 68 - 102 | Pass |
| TCHOL mg/dL | 155 - 225 | Pass | TCHOL mg/dL | 307 - 423 | Pass |
| TGL mg/dL | 60 - 136 | Pass | TGL mg/dL | 280 - 420 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

3/19/2024
Date