



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ Lactate Sensor  
Lot Number 24066033

Part Number 48553  
Exp Date 2026-03-06

Date of Manufacture 2024-03-06

UDI# (01)00385480485531(11)240306(17)260306(10)24066033

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	73 - 87	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

3/20/2024  
Date