

OI-922

Page 32 of 66

Rev. BJ

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, NOVA MAX PRO CREATININE eGFR, 50 COUNT
Catalog Number:	63983
Nova Biomedical P/N	63983
Control Level 1: CV (%) \leq 15% Control Level 2: CV (%) \leq 10%	PASS / FAIL PASS / FAIL
Qty:	378
Lot No.	2730223317
Mfg. Date:	2023-11-13
Expiration Date:	2024-11-13
Unique Device Identifier (UDI):	(01)20385480639839 (11) <u>231113</u> (17) <u>241113</u> (10) <u>2730223317</u>

Marc Brosme
Quality Control Manager

12/6/2023
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454