



CERTIFICATE OF ANALYSIS

Nova Lactate Plus Control Solution Level 1
Lot Number 3824078301

Ref Number 40815
Exp Date 2026-09-18
Date of Manufacture 2024-03-18
UDI# (01)00385480408158(11)240318(17)260918(10)3824078301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	1.3 - 1.5	1.3

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

3/22/2024
Date