



**CERTIFICATE OF ANALYSIS**

Nova Cleaning Solution  
Lot Number 24073011

Ref Number 11272  
Exp Date 2025-03-13  
Date of Manufacture 2024-03-13  
UDI# (01)00385480112727(11)240313(17)250313(10)24073011

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	1.21 - 1.51%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

3/25/2024  
Date