



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ Lactate Sensor  
Lot Number 24068021

Part Number 48553  
Exp Date 2026-03-08

Date of Manufacture 2024-03-08

UDI# (01)00385480485531(11)240308(17)260308(10)24068021

|             | <u>Spec</u>   | <u>Pass/Fail</u> |
|-------------|---------------|------------------|
| Slope       | 11.00 - 45.00 | Pass             |
| pHOx Ctl L3 | 73 - 87       | Pass             |
| Set A L1    | ≥ 470         | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

3/25/2024  
Date