



CERTIFICATE OF ANALYSIS

Bio pHox Cal Pack
Lot Number 24074080

Part Number 41874
Exp Date 2025-09

Std A	L/N	24064057			Std B	L/N	24064056			Std C	L/N	24061040		
		<u>Spec</u>		<u>Pass/Fail</u>			<u>Spec</u>		<u>Pass/Fail</u>			<u>Spec</u>		<u>Pass/Fail</u>
	Na mmol/L	139.0 - 141.0		Pass	HCO3 mmol/L	39.2 - 40.8		Pass		pH @ 37C	7.352 - 7.358		Pass	
	HCO3 mmol/L	19.6 - 20.4		Pass										
		<u>Spec</u>		<u>Pass/Fail</u>			<u>Spec</u>		<u>Pass/Fail</u>			<u>Spec</u>		<u>Pass/Fail</u>
Std D	L/N	24061041			Reference	L/N	24061036							
	pH @ 37C	6.802 - 6.808		Pass		KCl mol/L	1.8 - 2.2		Pass					
	Na mmol/L	69.5 - 70.5		Pass										

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO3, pH, KCl.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

3/27/2024
Date