

CERTIFICATE OF ANALYSIS

Performance Check Solution Lot Number 24071064 Ref Number 17457 Exp Date 2025-09-11 Date of Manufacture 2024-03-11 UDI# (01)00385480174572(11)240311(17)250911(10)24071064

	<u>Spec</u>	Pass/Fail
Glu mg/dL	689.5 - 710.5	Pass
BUN mg/dL	97.0 - 103.0	Pass
Creatinine mg/dL	1.54 - 1.66	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector Title

<u>3/27/2024</u> Date