



CERTIFICATE OF ANALYSIS

Stat Profile pHox/Basic Calibrator Cartridge C
Lot Number 24074079

Ref Number 23936
Exp Date 2025-09-01
Date of Manufacture 2024-03-01
UDI# (01)00385480239363(11)240301(17)250901(10)24074079

Cal A L/N 24064057			Cal B L/N 24064056			Cal C L/N 24061040		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HCO ₃ mmol/L	20.1 - 20.9	Pass	HCO ₃ mmol/L	40.2 - 41.8	Pass	pH @ 37C	7.342 - 7.348	Pass
Na mmol/L	139.3 - 141.3	Pass						
Cal D L/N 24061041			Reference L/N 24061036					
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>			
pH @ 37C	6.792 - 6.798	Pass	KCl mol/L	1.8 - 2.2	Pass			
Na mmol/L	71.5 - 72.5	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

03/28/2024
Date