



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS /FAIL PASS /FAIL PASS /FAIL
Qty:	8,028
Lot No.	4924072129
Mfg. Date:	2024-03-12
Expiration Date:	2025-03-12
Unique Device Identifier (UDI):	(01)20385480432720 (11) <u>240312</u> (17) <u>250312</u> (10) <u>4924072129</u>


Quality Control Manager

3/19/24
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454