



CERTIFICATE OF ANALYSIS

STP® pHOx Ultra™ / CCX™ Creatinine Sensor
Lot Number 24072022

Part Number 48595
Exp Date 2026-03-12
Date of Manufacture 2024-03-12
UDI# (01)00385480485951(11)240312(17)260312(10)24072022

	<u>Specs</u>	<u>Pass/Fail</u>
Slope	≥ 5.00	Pass
pHOx Ctl L4	0.7 - 1.3	Pass
pHOx Ctl L5	5.2 - 7.2	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Atwishi.M

Approval

QC Inspector

Title

03/28/2024

Date