



**CERTIFICATE OF ANALYSIS**

Nova Cleaning Solution  
Lot Number 22125083

Ref Number 11272  
Exp Date 2023-05-05  
Date of Manufacture 2022-05-05  
UDI# (01)00385480112727(11)220505(17)230505(10)22125083

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	1.21 - 1.51%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

5/20/2022  
Date