



**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, NOVA VET KETONE, 25 CT., PACKAGED
Catalog Number:	52587
Nova Biomedical P/N	52587
<u>Target Range</u> <u>MPE</u> <u>SD/CV (%)</u>	
0.1-0.8 mmol/L    +/- 0.3 mmol/L    ≤ 0.3 mmol/L	<u>PASS</u> / FAIL
1.5-2.5 mmol/L    +/- 20%              ≤ 15%	<u>PASS</u> / FAIL
3.5-5.0 mmol/L    +/- 20%              ≤ 15%	<u>PASS</u> / FAIL
Qty:	398
Lot No.	1760823219
Mfg. Date:	2023-08-07
Expiration Date:	2025-08-07
Unique Device Identifier (UDI):	(01)20385480 <u>525873</u> (11) <u>230807</u> (17) <u>250807</u> (10) <u>1760823219</u>

Maie Brosme  
Quality Control Manager

4/2/2024  
Date

Nova Biomedical Corporation      200 Prospect Street      Waltham, MA 02454