



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane
Lot Number 24059016

Part Number 48552
Exp Date 2025-05-28
Date of Manufacture 2024-02-28
UDI# (01)00385480485524(11)240228(17)250528(10)24059016

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	28.55
pHOx Ultra/CCX Ctl L4	73 - 87	79
pHOx Ultra/CCX Ctl L5	255 - 305	296

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/4/2024

Date