



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Linearity Kit
Lot Number 0524068990

Ref Number 42173

Exp Date 2026-09-08

Date of Manufacture 2024-03-08

UDI# (01)00385480421737(11)240308(17)260908(10)0524068990

Level 1 L/N	0524073301
Level 1 Exp Date	2026-09-13
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	16 - 22 20

Level 2 L/N	0524068302
Level 2 Exp Date	2026-09-08
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	57 - 65 63

Level 3 L/N	0524068303
Level 3 Exp Date	2026-09-08
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	105 - 125 113

Level 4 L/N	0524072304
Level 4 Exp Date	2026-09-12
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	275 - 325 301

Level 5 L/N	0524072305
Level 5 Exp Date	2026-09-12
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	470 - 540 506

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. Analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

4/4/2025

Date