



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 500 Sample
Lot Number 24081066

Ref Number 53464
Exp Date 2025-09-12
Date of Manufacture 2024-03-12
UDI# (01)00385480534642(11)240312(17)250912(10)24081066

Cal A	L/N	24074033		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.354	N/A	
	pCO2 mmHg	24.7	N/A	

Cal B	L/N	24074035		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	24072044		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	50.6	N/A	
	pO2 mmHg	99.6	N/A	

Reference	L/N	24074034		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCl	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

04/02/2024
Date