

CERTIFICATE OF ANALYSIS

Prime Cal Cartridge ES Comp 600 Sample Lot Number 24093095

D Sample Ref Number 53462 Exp Date 2025-09-18 Date of Manufacture 2024-03-18 UDI# (01)00385480534628(11)240318(17)250918(10)24093095

Cal A L/N 24078034		Cal E	Cal B L/N 24078035			Reference L/N 24079012		
<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail	
Na mmol/L 134.1 - 135.9	Pass	Na mmol/L	76.5 - 77.5	Pass	KCI mol/L	1.8 - 2.2	Pass	
K mmol/L 4.02 - 4.08	Pass	K mmol/L	9.80 - 10.20	Pass				
CI mmol/L 114.1 - 115.9	Pass	CI mmol/L	44.6 - 45.5	Pass				
iCa mmol/L 0.99 - 1.01	Pass	iCa mmol/L	1.98 - 2.02	Pass				
iMg mmol/L 0.40 - 0.50	Pass	iMg mmol/L	1.48 - 1.52	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Fint

Approval

QC Inspector Title

<u>04/08/2024</u> Date