CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, NOVA PRO GLUCOSE, PACKAGED 1800 COUNT
Catalog Number:	53881
Nova Biomedical P/N	53881
Target Range (mg/dL) SD(mg/dL) /CV (%) 30-50 ≤ 4 mg/dL 51-110 ≤ 6.0% 111-150 ≤ 5.0% 151-250 ≤ 5.0% 251-400 ≤ 5.0% Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method ≥ 95% Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method ≥ 98%	PASS / FAIL PASS / FAIL PASS / FAIL PASS / FAIL PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	1,680
Lot No.	1890424085
Mfg. Date:	2024-03-25
Expiration Date:	2026-03-25
Unique Device Identifier (UDI):	(01)20385480538811 (11)240325 (17)260325 (10)1890424085

Quality Control Manager Date 4/10/24

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

Released Confidential