



**CERTIFICATE OF ANALYSIS**

StatSensor Creatinine Control Solution Level 3  
Lot Number 5024059243

Ref Number 43923

Exp Date 2026-02-28

Date of Manufacture 2024-02-28

UDI# (01)00385480439237(11)240228(17)260228(10)5024059243

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	5.50 - 7.00	6.02

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/9/2024

Date