



CERTIFICATE OF ANALYSIS

Nova Cleaning Solution
Lot Number 22294060

Ref Number 11272
Exp Date 2023-10-21
Date of Manufacture 2022-10-21
UDI# (01)00385480112727(11)221021(17)231021(10)22294060

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	1.21 - 1.51%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

11/14/2022
Date