



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ BUN Sensor  
Lot Number 24085044

Part Number 48555  
Exp Date 2025-06-25  
Date of Manufacture 2024-03-25  
UDI# (01)00385480485555(11)240325(17)250625(10)24085044

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	$\geq 9.60$	Pass
pHOx Ctl L4	14 - 24	Pass
pHOx Ctl L5	49 - 63	Pass
Std D mV	$0.00 \pm 60.00$	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/11/2024

Date