



**CERTIFICATE OF ANALYSIS**

Allegro UACR Test Cartridge  
Lot Number 24100014

Ref Number 54214  
Exp Date 2025-04-09  
Date of Manufacture 2024-04-09  
UDI# (01)00385480542142(11)240409(17)250409(10)24100014

	Ctl Level 1			Ctl Level 2	
	Spec	Pass/Fail		Spec	Pass/Fail
Albumin mg/L	31 - 55	Pass	Albumin mg/L	115 - 191	Pass
Creatinine mg/dL	67 - 111	Pass	Creatinine mg/dL	161 - 277	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

4/12/2024  
Date