



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Lactate Sensor
Lot Number 24086008

Part Number 48553
Exp Date 2026-03-26

Date of Manufacture 2024-03-26

UDI# (01)00385480485531(11)240326(17)260326(10)24086008

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------------|---------------|------------------|
| Slope | 11.00 - 45.00 | Pass |
| pHOx Ctl L3 | 78 - 92 | Pass |
| Set A L1 | ≥ 470 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

4/15/2024
Date