

CERTIFICATE OF ANALYSIS

STAT PROFILE® pHOx Ultra™/CCX™ pCO2 Sensor Lot Number 24086013 Ref Number 48600 Exp Date 2026-03-26 Date of Manufacture 2024-03-26

UDI# (01)00385480486002(11)240326(17)260326(10)24086013

 Spec
 Pass/Fail

 Slope
 9.40 - 11.60
 Pass

 Std C mV
 0.00 ± 150.00
 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/15/2024

Date