



**CERTIFICATE OF ANALYSIS**

Allegro Control 2 UACR  
Lot Number 8524086242

Exp Date 2026-03-26  
Date of Manufacture 2024-03-26  
Ref Number 51031  
UDI# (01)00385480510318(11)240326(17)260326(10)8524086242

	Spec	Pass/Fail
Albumin, mg/dL	127-193	Pass
Creatinine, mg/dL	196-288	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

4/15/2024  
Date