



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA		
Description:	TEST STRIPS, STATSTRIP™ KETONE, PACKAGED 50 CT		
Catalog Number:	46951		
Nova Biomedical P/N	46951		
<u>Target Range</u>	<u>MPE</u>	<u>SD/CV (%)</u>	
0.1-0.8 mmol/L	+/- 0.3 mmol/L	≤ 0.3 mmol/L	<u>PASS</u> / FAIL
1.5-2.5 mmol/L	+/- 20%	≤ 15%	<u>PASS</u> / FAIL
3.5-5.0 mmol/L	+/- 20%	≤ 15%	<u>PASS</u> / FAIL
Qty:	21,064		
Lot No.	2224099249		
Mfg. Date:	2024-04-08		
Expiration Date:	2026-04-08		
Unique Device Identifier (UDI):	(01)20385480 <u>469511</u> (11) <u>240408</u> (17) <u>260408</u> (10) <u>2224099249</u>		

Marie Broome
 Quality Control Manager

4/18/2024
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454