



CERTIFICATE OF ANALYSIS

BioPro Flex Deproteinizing Sol'n Cartridge
Lot Number 24068038

Ref Number 45192
Exp Date 2025-03

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl %	0.70 - 0.90	Pass
pH	9.00 - 11.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

4/17/2024

Date