



CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor
Lot Number 24102012

Part Number 48591
Exp Date 2027-04-11
Date of manufacture 2024-04-11
UDI#(01)00385480485913(11)240411(17)270411(10)24102012

	<u>Spec</u>	<u>Mean Result</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	9.36	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

4/17/2024
Date