



CERTIFICATE OF ANALYSIS

Nova 8 Calibrator Pack  
Lot Number 410721

Ref Number 15196  
Exp Date 2025-08-29  
Date of Manufacture 2024-02-29  
UDI#(01)00385480151962(11)240229(17)250829(10)410721

Cal A L/N 24061005		
	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.0 - 141.0	Pass
K mmol/L	3.97 - 4.03	Pass
iCa mmol/L	0.99 - 1.01	Pass
iMg mmol/L	0.495 - 0.505	Pass
pH @ 37C	7.274 - 7.294	Pass

Cal B L/N 24061008		
	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	74.5 - 75.5	Pass
K mmol/L	9.9 - 10.1	Pass
iCa mmol/L	1.98 - 2.02	Pass
iMg mmol/L	0.495 - 0.505	Pass
pH @ 37C	6.730 - 6.750	Pass

Cal C L/N 24073020		
	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.0 - 141.0	Pass
K mmol/L	3.97 - 4.03	Pass
iCa mmol/L	0.99 - 1.01	Pass
iMg mmol/L	1.49 - 1.52	Pass
pH @ 37C	7.374 - 7.394	Pass

Reference L/N 24060052		
	<u>Spec</u>	<u>Pass/Fail</u>
KCl mol/L	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

*Auishi M*

Approval

QC Inspector  
Title

4/18/2024  
Date