



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Lactate Sensor
Lot Number 24103025

Part Number 48553
Exp Date 2026-04-12

Date of Manufacture 2024-04-12

UDI# (01)00385480485531(11)240412(17)260412(10)24103025

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	78 - 92	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/26/2024

Date