



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 4
Lot Number 0423047304

Ref Number 50539
Exp Date 2025-08-16

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	555 - 625	592

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/20/2023

Date