



**CERTIFICATE OF ANALYSIS**

Nova 4/10/11/13 Lithium Sensor  
Lot Number 24093083

Ref Number 15742  
Exp Date 2025-04-02

Date of Manufacture 2024-04-02

UDI# (01)10385480157428(11)240402(17)250402(10)24093083

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	10.3 - 11.7	Pass
Chem Ctl L 2 mmol/L	1.77 - 2.27	Pass
Set A L 2 mmol/L	1.96 - 2.36	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

<sup>2</sup>Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/26/2024

Date