



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample
Lot Number 24109058

Ref Number 53360
Exp Date 2025-10-11
Date of Manufacture 2024-04-11
UDI# (01)00385480533607(11)240411(17)251011(10)24109058

| | | | | |
|-------|-----------|------------------------|---------------|--|
| Cal A | L/N | 24103001 | | |
| | | <u>Assigned Value*</u> | <u>Result</u> | |
| | pH | 7.353 | N/A | |
| | pCO2 mmHg | 25.0 | N/A | |

| | | | | |
|-------|----------|---------------|------------------|--|
| Cal B | L/N | 24103002 | | |
| | | <u>Spec</u> | <u>Pass/Fail</u> | |
| | pH | 6.837 - 6.843 | Pass | |
| | pO2 mmHg | 167.0 | N/A | |

| | | | | |
|-------|-----------|------------------------|---------------|--|
| Flush | L/N | 24102025 | | |
| | | <u>Assigned Value*</u> | <u>Result</u> | |
| | pCO2 mmHg | 52.5 | N/A | |
| | pO2 mmHg | 93.7 | N/A | |

| | | | | |
|-----------|-----|-------------|------------------|--|
| Reference | L/N | 24103011 | | |
| | | <u>Spec</u> | <u>Pass/Fail</u> | |
| | KCl | 1.8 - 2.2 | Pass | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

05/01/2024
Date