



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA		
Description:	TEST STRIPS, STATSTRIP LACTATE, PACKAGED		
Catalog Number:	47486		
Nova Biomedical P/N	47486		
<u>Target Range</u>	<u>CV (%)</u>	<u>SD (mmol/L)</u>	
0.8-2.5 mmol/L	≤ 6.5%	±0.3	<u>PASS</u> / FAIL
3.0-5.0 mmol/L	≤ 5.5%	±0.5	<u>PASS</u> / FAIL
10-14 mmol/L	≤ 5.0%	±1.2	<u>PASS</u> / FAIL
Qty:	7,510		
Lot No.	2624120249		
Mfg. Date:	2024-04-29		
Expiration Date:	2026-04-29		
Unique Device Identifier (UDI):	(01)20385480 <u>474867</u> (11) <u>240429</u> (17) <u>260429</u> (10) <u>2624120249</u>		

Marc Brone
 Quality Control Manager

5/3/2024
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454