



**CERTIFICATE OF ANALYSIS**

Nipro Glucose and  $\beta$ -Ketone Control Solution Level 1  
Lot Number 0424078301

Ref Number 11-115  
Exp Date 2026.09.30

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	60
Ketone mmol/L	0.1 - 0.2	0.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi M*

Approval

QC Inspector

Title

5/6/2024

Date