



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHOx Ultra™/CCX™ Potassium Sensor
Lot Number 24115059

Ref Number 48582

Exp Date 2025-07-24

Date of Manufacture 2024-04-24

UDI# (01)00385480485821(11)240424(17)250724(10)24115059

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 to 11.00	Pass
Std C mV	0 ± 50.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

5/10/2024

Date