



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA		
Description:	TEST STRIPS, STATSTRIP™ KETONE, PACKAGED 50 CT		
Catalog Number:	46951		
Nova Biomedical P/N	410951		
<u>Target Range</u>	<u>MPE</u>	<u>SD/CV (%)</u>	
0.1-0.8 mmol/L	+/- 0.3 mmol/L	≤ 0.3 mmol/L	PASS / FAIL
1.5-2.5 mmol/L	+/- 20%	≤ 15%	PASS / FAIL
3.5-5.0 mmol/L	+/- 20%	≤ 15%	PASS / FAIL
Qty:	14,202		
Lot No.	2224127249		
Mfg. Date:	2024-05-06		
Expiration Date:	2026-05-06		
Unique Device Identifier (UDI):	(01)20385480 469511 (11) 240506 (17) 260506 (10) 2224127249		

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454