



CERTIFICATE OF ANALYSIS
StatStrip Glucose Control Solution Level 3

Nova Ref Number 54522
LifeScan Ref Number 02315602
Lot Number 0424099243

Date of Manufacture 2024-04-08
Exp Date 2026-04-08
UDI # (01)00385480545228(11)240408(17)260408(10)0424099243

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	298

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

05/13/2024
Date