

CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamate Membrane Kit Lot Number 24127004

Part Number 42484 Exp Date 2024-11

| | Spec* | Mean Result |
|---------------|---------------|-------------|
| Slope | 10.00 - 60.00 | 37.17 |
| Biopro Ctl L4 | 0.27 - 0.77 | 0.46 |
| Biopro Ctl L5 | 1.78 - 2.38 | 2.10 |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

<u>QC Inspector</u> Title

<u>5/13/2024</u> Date