



CERTIFICATE OF ANALYSIS

Bioprofile 300 Lactate Membrane
Lot Number 24128015

Part Number 34573
Exp Date 2025-05

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.0 - 60.0	43.32
Biopro Ctl L 6	0.75 - 1.25	1.08
Biopro Ctl L 7	2.50 - 3.50	3.09

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

5/16/2024
Date