

## **CERTIFICATE OF ANALYSIS**

Allegro HbA1c Test Cartridge Lot Number 24135006

tridge Ref Number 54215 Exp Date 2025-05-14 Date of Manufacture 2024-05-14 UDI# (01)00385480542159(11)240514(17)250514(10)24135006

	Ctl Level 1			Ctl Level 2	
	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail
HbA1c %	4.9 - 6.7	Pass	HbA1c %	8.1 - 11.7	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Muyht

Approval

<u>QC Inspector</u> Title

<u>5/17/2024</u> Date