



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100 Plus Glucose Membrane
Lot Number 24136034

Part Number 24458
Exp Date 2025-08

	<u>Spec*</u>	<u>Mean Result</u>
Slope	12.0 - 60.0	34.0
Biopro Ctl L1	0.75 - 1.25	1.09
Biopro Ctl L3	12.75 - 17.25	15.50

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector

Title

5/17/2024

Date