



CERTIFICATE OF ANALYSIS
Nova StatStrip Glucose Control Solution Level 1

Nova Reference 54520
LifeScan Reference 02315402
Lot Number 0424095241

Date of Manufacture 2024-04-04
Exp Date 2026-04-04
UDI# (01)00385480545204(11)240404(17)260404(10)0424095241

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	62

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

05/17/2024
Date