



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 24134001

Ref Number 54652
Exp Date 2025-11-13
Date of Manufacture 2024-05-13
UDI# (01)00385480546522(11)240513(17)251113(10)24134001

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HDL mg/dL	28 - 50	Pass	HDL mg/dL	68 - 102	Pass
TCHOL mg/dL	155 - 225	Pass	TCHOL mg/dL	307 - 423	Pass
TGL mg/dL	60 - 136	Pass	TGL mg/dL	280 - 420	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

5/20/2024
Date