



CERTIFICATE OF ANALYSIS

Nova Na/pH Conditioning Solution
Lot Number 24123044

Ref Number 00018
Exp Date 2025-08-02
Date of Manufacture 2024-05-02
UDI# (01)00385480000185(11)240502(17)250802(10)24123044

	<u>Spec</u>	<u>Pass/Fail</u>
pH	≤ 5.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi M

Approval

QC Inspector

Title

5/21/2024

Date