



**CERTIFICATE OF ANALYSIS**

STAT PROFILE® pHox Ultra™/CCX™ pO2 Sensor  
Lot Number 24134044

Ref Number 48598  
Exp Date 2026-05-13  
Date of Manufacture 2024-05-13  
UDI# (01)00385480485982(11)240513(17)260513(10)24134044

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

5/24/2024  
Date