



CERTIFICATE OF ANALYSIS

Preheater Deproteinizing Sol'n
Lot Number 22140047

Ref Number 12704
Exp Date 2023-05-20
Date of Manufacture 2022-05-20
UDI# (01)00385480127042(11)220520(17)230520(10)22140047

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.70 - 0.90%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Awishi M

Approval

QC Inspector
Title

6/3/2022
Date