



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Lactate Linearity Solutions Levels 1 - 4  
Lot Number 2824117990

Ref Number 47552  
Exp Date 2026-10-26  
Date of Manufacture 2024-04-26  
UDI# (01)00385480475525(11)240426(17)261026(10)2824117990

Level 1 L/N	2824117301		Level 2 L/N	2824117302	
Level 1 Exp Date	2026-10-26		Level 2 Exp Date	2026-10-26	
	Spec	Mean Result		Spec	Mean Result
Lactate mmol/L	0.5 - 0.7	0.6	Lactate mmol/L	2.2 - 2.6	2.4
Level 3 L/N	2824120303		Level 4 L/N	2824121304	
Level 3 Exp Date	2026-10-29		Level 3 Exp Date	2026-10-30	
	Spec	Mean Result		Spec	Mean Result
Lactate mmol/L	5.9 - 6.7	6.2	Lactate mmol/L	15.5 - 18.0	17.3

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

5/31/2024  
Date